

2400S Ophthalmic Ultrasound



2400S A/B Scan with normal, vitreous body enhancement, retina observation mode, mainly used for diagnosis of intraocular diseases, display the location, shape range of the focus of infection and the relationship with the surrounding tissue. Can be diagnosed vitreous opacity, retinal detachment, eye base tumors etc. eye diseases. A scan is used to measure anterior chamber depth, lens thickness, axial length, calculate diopter of implant IOL as well.

B scan:

Frequency: 10MHz/20MHz (optional) ,Magnetic driven, noiseless

Scanning Mode: Sector Scanning

Magnify: Multi continuous magnification,Real-Time magnification

Resolution: Lateral $\leq 0.3\text{mm}$; Vertical $\leq 0.2\text{mm}$

Geometry position precision: Lateral $\leq 10\%$; Vertical $\leq 5\%$

Depth:60mm

Enhance the part of vitreous body and retina

Gain of probe: 30dB-105dB

Scanning Angle:53°

Gray Scale: 256

False Color: Multi colors. OCT

measurement type: multigroup distances, perimeters and areas

Image postprocessing: multiple curves processing, Pseudo-color processing curve

Movies: 100 images movie review, AVI JPG format image output

Add: Block A, Lotus Industrial Park, Economic Development Zone, Hefei City, Anhui, China.

Tel: 0086-551-63802963

Fax: 0086-551-63802920

Web:www.ari-cn.com

Email: sales@ari-cn.com

A scan:

Frequency:10MHz, with LED

Depth: 40mm

Precision:±0.05 mm

Measurement: Anterior chamber depth, lens thickness, vitreous body length, total length and average

Eye mode: Phakic / Aphakic / Dense / Various IOL

IOL Formula: SRK-II, SRK-T, HOFFER-Q, HOLLADAY, BINKHORST-II, HAIGIS

Stat. Calculation: Average and standard deviation

Store: 10 Scanning results for each eye

Others:

Display Mode :B, B+B, B+A, A

Hint: preset keyword

Case Search: Multi-keywords

Screen: 15 inch LCD

Built-in battery: can use 4 hours

User-defined report template poor credit secured loans

